


Calibrate Thermometers Often

Make sure temperatures are correct.


Ice Water Method

1. Fill a big glass with ice.
2. Add clean tap water until glass is full.
3. Stir well.
4. Put thermometer stem into ice water. The entire sensing area needs to be under water.
5. Wait 30 seconds.
6. Hold the adjusting nut with a tool and twist the head of the thermometer until it reads 32°F.
7. Wait 30 seconds and recheck temperature.



Boiling Water Method

1. Bring clean tap water to a boil in a deep pan.
2. Put thermometer stem into the boiling water. The entire sensing area needs to be under water.
3. Wait 30 seconds.
4. Hold the adjusting nut with a tool and twist the head of the thermometer until it reads 212°F (100°C) or the appropriate boiling temperature.
5. Wait 30 seconds and recheck temperature.



NOTES:

- Do not let the thermometer touch the bottom or sides of the glass or pan.
- The thermometer stem must remain in the water.
- Press the reset button (if available) on a digital thermometer to adjust the readout.